

Date: January 5, 2010

Subject: Marfs, Bodily Fluids Resistance ASTM F 1862

Background: ASTM F 1862 test has been conducted to evaluate the resistance of Marfs to a fixed amount of bodily fluids at a high velocity. This is a pass/fail determination based on the penetration or not of the bodily fluid.

Test and Test Results:

ASTM F 1862	Result
Marfs	Pass

Conclusion: Marfs have demonstrated to provide an adequate amount of protection against the spread and transfer of bodily fluids. Marfs can be used to help prevent the spread of infectious diseases.

